

ExionLC 2.0 Valve Drive model DR-200 for use with models ExionLC 2.0 Valve 2 Position 6 Port and ExionLC 2.0 Valve 8 Port Multi-Position

UK DECLARATION OF CONFORMITY

5087274

| SCIEX | (According to BS EN ISO/IEC 17050-1) | | |
|--|---|--|--|
| We | AB SCIEX Pte. Ltd | | |
| of | Blk 33, #04-06, Marsiling Ind. Estate Road 3, Woodlands Central Industrial Estate, Singapore 739256 | | |
| declare under sole responsibility that the product as originally delivered | | | |
| Equipment: Product Name: | Liquid Chromatography Valve Drive ExionLC 2.0 Valve Drive model DR-200 (for use with ExionLC 2.0 Valve 2 | | |

complies with the essential requirements of the following applicable UK legislations, and carries the marking accordingly:

Position 6 Port and ExionLC 2.0 Valve 8 Port Multi-Position)

- Electrical Equipment (Safety) Regulations 2016
- Electromagnetic Compatibility Regulations 2016

Valve Drive

- Waste Electrical and Electronic Equipment Regulations 2013
- The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 and applicable amendments

The product was tested in a typical configuration and conforms with the following product standards:

| Category | Standard | Classification |
|----------|---|----------------|
| Safety | BS EN 61010-1:2010 +A1:2019 BS EN 61010-2-081:2020 | Class I |
| EMC | BS EN 61326-1:2013 | Class B |

I, the undersigned, hereby declare that the equipment specified above conforms to the above Regulations and Standards.

7/28/2022

X Nichole Riek

Product Options:

Signed by: Riek, Nichole Nichole Riek Vice President, Quality Assurance and Regulatory Affairs