

SCIEX ExionLC 2.0 Binary Pump⁺ model BP-200⁺ and SCIEX ExionLC 2.0 LPG Pump⁺ model LPGP-200⁺

DECLARATION OF CONFORMITY

5302197

F1670-10 Rev C D5304654 (1/2)



DECLARATION OF CONFORMITY

(According to EN ISO/IEC 17050-1)



We AB SCIEX Pte. Ltd

of Blk 33, #04-06, Marsiling Ind. Estate Road 3, Woodlands Central

Industrial Estate, Singapore 739256

declare under sole responsibility that the product as originally delivered

Equipment: Liquid Chromatography Pump

Product Name: SCIEX ExionLC 2.0 Binary Pump* model BP-200* and SCIEX ExionLC 2.0 LPG Pump*

model LPGP-200+

Equipment Components: Pumps

EU:

complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:

2014/35/EU - Low Voltage Directive

2014/30/EU - Electromagnetic Compatibility Directive

2012/19/EU - Waste Electrical and Electronic Equipment Directive 2011/65/EU and 2015/863/EU - Restriction of Hazardous Substances Directive

Australia and New Zealand:

complies with the requirements of the relevant ACMA Standards made under the *Radiocommunications Act 1992* and the *Telecommunications Act 1997* as well as the New Zealand Radiocommunications Act 1989.

The product was tested in a typical configuration and conforms with the following standards:

Category Standard Classification

Safety EU: IEC 61010-1:2010+ Amd 1: 2016 / EN Equipment Class I, Pollution Degree 2

61010-1:2010 + Amd 1: 2019 IEC 61010-2-081:2019 / EN 61010-2-

081:2020

EMC Limit
EU: IEC 61326-1:2012 / EN 61326-1:2013 Class B Group 1 ISM

: IEC 61326-1:2012 / EN 61326-1:2013 EN 61000-3-2: 2015-03

EN 61000-3-3: 2014-03

Australia/New AS CISPR 11:2017

 Zealand:
 ICES-001 (CSA CISPR 11:2019)
 Class B

 USA:
 FCC Part 15
 Class B

I, the undersigned, hereby declare that the equipment specified above conforms to the above Directives, Regulations and Standards.

Class B

7/25/2022

X Nichole Riek

Signed by: Riek, Nichole
Nichole Riek
Vice President,
Quality Assurance and
Regulatory Affairs